

Lab gas cylinder acceptance checklist

Date and time of inspection: _____ **Signature of laboratory inspection:** _____

No.	Inspection content		Inspection results		Remarks
			Does it match Yes/No		
1.	Gas types meets lab's requirements				
2.	Gas purity meets the requirements				
3.	Gas cylinder pressure meets the requirements				
4.	Gas cylinder safety label is in place				
5.	Cylinder accessories are complete and in good condition	Valve protection cover			
		Valve protection plug			
		Angle valve 's handwheel			
6.	No obvious defects in the appearance of the cylinders				
7.	Leak test after connection of gas cylinder				
8.	Gas cylinders take fixed measures				

Remark.

- If the inspection results of any of the items 1 to 7 are unqualified, the cylinder is considered unqualified cylinders cannot be accepted.
- Item 8 If this situation exists, please contact Campus Operations Department immediately for gas cylinder fixing.

Gas cylinders are reasonably fixed in the transport vehicle and not easily displaced



Cylinder accessories are complete and in good condition (Valve protection cover)





Central gas supply cylinders must be covered with a cap to protect the cylinder valve



Cylinder accessories are complete and in good condition (Valve protection plug)



Cylinder accessories are complete and in good condition (Angle valve handwheel)

国标无缝钢瓶

一瓶一合格证

- 1 拒绝翻新瓶
- 2 寿命达30年
- 3 GB5099标准



No obvious defects in the appearance of the cylinders



Equipped with gas cylinder transportation tools (gas cylinder special cart)



Gas cylinder safety label complete